

CLIENT DETAILS

Contact **John Fox**  
 Client **NAMOI COTTON CO-OPERATIVE LTD**  
 Address **1B KITCHENER STREET  
 TOOWOOMBA QLD 4350**

Telephone **0429 903 079**  
 Facsimile **61 7 46316184**  
 Email **jfox@namoicotton.com.au**

Project **Mungindi Discharge Event Monitoring**  
 Order Number **(Not specified)**  
 Samples **3**  
 Date Started **12 Jan 2015**

LABORATORY DETAILS

Manager **Andrew Tomlins**  
 Laboratory **SGS Brisbane Environmental**  
 Address **59 Bancroft Road  
 PINKENBA QLD 4008**

Telephone **+61 7 3622 4700**  
 Facsimile **+61 7 3622 4799**  
 Email **au.environmental.brisbane@sgs.com**

SGS Reference **BE011457 R0**  
 Report Number **0000035639**  
 Date Reported **16 Jan 2015**  
 Date Received **09 Jan 2015**

COMMENTS

Accredited for compliance with ISO/IEC 17025. NATA accredited laboratory 2562(20707/1706).

NO3, NO2, TN and TP subcontracted to SGS Sydney, Unit 16 33 Maddox St Alexandria NSW 2015, NATA Accreditation Number: 2562, Site Number: 4354, SE135022 R0.

SIGNATORIES



Caroline McDermid  
Inorganics Supervisor



Michael Morrison  
Senior Organic Chemist

	Sample Number	BE011457.001	BE011457.002	BE011457.003	
	Sample Matrix	Water	Water	Water	
	Sample Date	05 Jan 2015	05 Jan 2015	05 Jan 2015	
	Sample Name	MG Site 3	MG Site 4	MG Site 5	
	Drainage Channel	Sedimentation	Sedimentation	Sedimentation	
Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module

pH in water Method: AN101

pH**	pH Units	0.1	6.7	6.6	7.3
------	----------	-----	-----	-----	-----

Conductivity and TDS by Calculation - Water Method: AN106

Conductivity @ 25 C	µS/cm	5	110	43	230
---------------------	-------	---	-----	----	-----

Total and Volatile Suspended Solids (TSS / VSS) Method: AN114

Total Suspended Solids Dried at 103-105°C	mg/L	1	380	720	120
---	------	---	-----	-----	-----

TKN Kjeldahl Digestion by Discrete Analyser Method: AN281/AN292

Total Nitrogen (calc)	mg/L	0.05	1.5	1.8	2.1
-----------------------	------	------	-----	-----	-----

Total Phosphorus by Kjeldahl Digestion DA in Water Method: AN279/AN293

Total Phosphorus (Kjeldahl Digestion)	mg/L	0.01	0.65	0.74	0.86
---------------------------------------	------	------	------	------	------

OC Pesticides in Water Method: AN400/AN420

Alpha BHC	µg/L	0.1	<0.1	<0.1	<0.1
Hexachlorobenzene (HCB)	µg/L	0.1	<0.1	<0.1	<0.1
Beta BHC	µg/L	0.1	<0.1	<0.1	<0.1
Lindane (gamma BHC)	µg/L	0.1	<0.1	<0.1	<0.1
Delta BHC	µg/L	0.1	<0.1	<0.1	<0.1
Heptachlor	µg/L	0.1	<0.1	<0.1	<0.1
Aldrin	µg/L	0.1	<0.1	<0.1	<0.1
Heptachlor epoxide	µg/L	0.1	<0.1	<0.1	<0.1
Isodrin	µg/L	0.1	<0.1	<0.1	<0.1
Gamma Chlordane	µg/L	0.1	<0.1	<0.1	<0.1
Alpha Chlordane	µg/L	0.1	<0.1	<0.1	<0.1
Alpha Endosulfan	µg/L	0.1	<0.1	<0.1	<0.1
p,p'-DDE	µg/L	0.1	<0.1	<0.1	<0.1
Dieldrin	µg/L	0.1	<0.1	<0.1	<0.1
Endrin	µg/L	0.1	<0.1	<0.1	<0.1
Beta Endosulfan	µg/L	0.1	<0.1	<0.1	<0.1
p,p'-DDD	µg/L	0.1	<0.1	<0.1	<0.1
Endosulfan sulphate	µg/L	0.1	<0.1	<0.1	<0.1
p,p'-DDT	µg/L	0.1	<0.1	<0.1	<0.1
Endrin ketone	µg/L	0.1	<0.1	<0.1	<0.1
Methoxychlor	µg/L	0.1	<0.1	<0.1	<0.1
Mirex	µg/L	0.1	<0.1	<0.1	<0.1

Parameter	Units	LOR	Drainage Channel NE of N Module	Sedimentation Pond E Module	Sedimentation Pond W Module
Sample Number			BE011457.001	BE011457.002	BE011457.003
Sample Matrix			Water	Water	Water
Sample Date			05 Jan 2015	05 Jan 2015	05 Jan 2015
Sample Name			MG Site 3	MG Site 4	MG Site 5

**OC Pesticides in Water** Method: AN400/AN420 (continued)  
Surrogates

Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module
d14-p-terphenyl (Surrogate)	%	-	78	60	66
2-fluorobiphenyl (Surrogate)	%	-	72	62	66
d5-nitrobenzene (Surrogate)	%	-	76	52	64

**OP Pesticides in Water** Method: AN400/AN420

Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module
Dichlorvos	µg/L	1	<1	<1	<1
Dimethoate	µg/L	1	<1	<1	<1
Diazinon (Dimpylate)	µg/L	0.5	<0.5	<0.5	<0.5
Fenitrothion	µg/L	0.2	<0.2	<0.2	<0.2
Malathion	µg/L	0.2	<0.2	<0.2	<0.2
Chlorpyrifos (Chlorpyrifos Ethyl)	µg/L	0.2	<0.2	<0.2	<0.2
Parathion-ethyl (Parathion)	µg/L	0.2	<0.2	<0.2	<0.2
Bromophos Ethyl	µg/L	0.2	<0.2	<0.2	<0.2
Methidathion	µg/L	0.5	<0.5	<0.5	<0.5
Ethion	µg/L	0.2	<0.2	<0.2	<0.2
Azinphos-methyl	µg/L	0.2	<0.2	<0.2	<0.2

Surrogates

Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module
d14-p-terphenyl (Surrogate)	%	-	78	60	66
d5-nitrobenzene (Surrogate)	%	-	72	62	66
2-fluorobiphenyl (Surrogate)	%	-	76	52	64

**Other SVOC Analytes in Water** Method: AN420

OPs

Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module
Carbophenothion	µg/L	0.5	<0.5	<0.5	<0.5
Chlorpyrifos-methyl	µg/L	0.5	<0.5	<0.5	<0.5
Dichlofenthion	µg/L	0.5	<0.5	<0.5	<0.5
Dioxathion	µg/L	2	<2	<2	<2
Famphur (Famphos)	µg/L	0.5	<0.5	<0.5	<0.5
Fonophos	µg/L	0.5	<0.5	<0.5	<0.5
Terbufos	µg/L	0.5	<0.5	<0.5	<0.5



ANALYTICAL REPORT

BE011457 R0

	Sample Number	BE011457.001	BE011457.002	BE011457.003	
	Sample Matrix	Water	Water	Water	
	Sample Date	05 Jan 2015	05 Jan 2015	05 Jan 2015	
	Sample Name	MG Site 3	MG Site 4	MG Site 5	
		Drainage Channel	Sedimentation	Sedimentation	
Parameter	Units	LOR	NE of N Module	Pond E Module	Pond W Module

Other SVOC Analytes in Water Method: AN420 (continued)  
other SVOCs

Thionazin	µg/L	1	<1	<1	<1
-----------	------	---	----	----	----

MB blank results are compared to the Limit of Reporting

LCS and MS spike recoveries are measured as the percentage of analyte recovered from the sample compared the the amount of analyte spiked into the sample.

DUP and MSD relative percent differences are measured against their original counterpart samples according to the formula : *the absolute difference of the two results divided by the average of the two results as a percentage*. Where the DUP RPD is 'NA' , the results are less than the LOR and thus the RPD is not applicable.

Conductivity and TDS by Calculation - Water Method: ME-(AU)-[ENV]AN106

Parameter	QC Reference	Units	LOR	MB	DUP %RPD	LCS %Recovery
Conductivity @ 25 C	LB017302	µS/cm	5	<5	0 - 1%	108%

OC Pesticides in Water Method: ME-(AU)-[ENV]AN400/AN420

Parameter	QC Reference	Units	LOR	MB	LCS %Recovery
Alpha BHC	LB017324	µg/L	0.1	<0.1	
Hexachlorobenzene (HCB)	LB017324	µg/L	0.1	<0.1	
Beta BHC	LB017324	µg/L	0.1	<0.1	
Lindane (gamma BHC)	LB017324	µg/L	0.1	<0.1	87%
Delta BHC	LB017324	µg/L	0.1	<0.1	
Heptachlor	LB017324	µg/L	0.1	<0.1	82%
Aldrin	LB017324	µg/L	0.1	<0.1	97%
Heptachlor epoxide	LB017324	µg/L	0.1	<0.1	
Isodrin	LB017324	µg/L	0.1	<0.1	86%
Gamma Chlordane	LB017324	µg/L	0.1	<0.1	86%
Alpha Chlordane	LB017324	µg/L	0.1	<0.1	
Alpha Endosulfan	LB017324	µg/L	0.1	<0.1	
p,p'-DDE	LB017324	µg/L	0.1	<0.1	85%
Dieldrin	LB017324	µg/L	0.1	<0.1	85%
Endrin	LB017324	µg/L	0.1	<0.1	89%
Beta Endosulfan	LB017324	µg/L	0.1	<0.1	
p,p'-DDD	LB017324	µg/L	0.1	<0.1	
Endosulfan sulphate	LB017324	µg/L	0.1	<0.1	
p,p'-DDT	LB017324	µg/L	0.1	<0.1	
Endrin ketone	LB017324	µg/L	0.1	<0.1	
Methoxychlor	LB017324	µg/L	0.1	<0.1	
Mirex	LB017324	µg/L	0.1	<0.1	90%

Surrogates

Parameter	QC Reference	Units	LOR	MB	LCS %Recovery
d14-p-terphenyl (Surrogate)	LB017324	%	-	90%	98%
2-fluorobiphenyl (Surrogate)	LB017324	%	-	88%	100%
d5-nitrobenzene (Surrogate)	LB017324	%	-	92%	108%

MB blank results are compared to the Limit of Reporting

LCS and MS spike recoveries are measured as the percentage of analyte recovered from the sample compared to the amount of analyte spiked into the sample. DUP and MSD relative percent differences are measured against their original counterpart samples according to the formula : *the absolute difference of the two results divided by the average of the two results as a percentage*. Where the DUP RPD is 'NA', the results are less than the LOR and thus the RPD is not applicable.

OP Pesticides in Water Method: ME-(AU)-[ENV]AN400/AN420

Parameter	QC Reference	Units	LOR	MB	LCS % Recovery
Dichlorvos	LB017324	µg/L	1	<1	
Dimethoate	LB017324	µg/L	1	<1	
Diazinon (Dimpylate)	LB017324	µg/L	0.5	<0.5	87%
Fenitrothion	LB017324	µg/L	0.2	<0.2	
Malathion	LB017324	µg/L	0.2	<0.2	
Chlorpyrifos (Chlorpyrifos Ethyl)	LB017324	µg/L	0.2	<0.2	80%
Parathion-ethyl (Parathion)	LB017324	µg/L	0.2	<0.2	67%
Bromophos Ethyl	LB017324	µg/L	0.2	<0.2	
Methidathion	LB017324	µg/L	0.5	<0.5	86%
Ethion	LB017324	µg/L	0.2	<0.2	
Azinphos-methyl	LB017324	µg/L	0.2	<0.2	

Surrogates

Parameter	QC Reference	Units	LOR	MB	LCS % Recovery
d14-p-terphenyl (Surrogate)	LB017324	%	-	90%	98%
d5-nitrobenzene (Surrogate)	LB017324	%	-	88%	100%
2-fluorobiphenyl (Surrogate)	LB017324	%	-	92%	108%

Other SVOC Analytes in Water Method: ME-(AU)-[ENV]AN420

OPs

Parameter	QC Reference	Units	LOR	MB
Carbophenothion	LB017324	µg/L	0.5	<0.5
Chlorpyrifos-methyl	LB017324	µg/L	0.5	<0.5
Dichlofenthion	LB017324	µg/L	0.5	<0.5
Dioxathion	LB017324	µg/L	2	<2
Famphur (Famophos)	LB017324	µg/L	0.5	<0.5
Fonophos	LB017324	µg/L	0.5	<0.5
Terbufos	LB017324	µg/L	0.5	<0.5

other SVOCs

Parameter	QC Reference	Units	LOR	MB
Thionazin	LB017324	µg/L	1	<1

MB blank results are compared to the Limit of Reporting  
 LCS and MS spike recoveries are measured as the percentage of analyte recovered from the sample compared the the amount of analyte spiked into the sample.  
 DUP and MSD relative percent differences are measured against their original counterpart samples according to the formula : *the absolute difference of the two results divided by the average of the two results as a percentage*. Where the DUP RPD is 'NA' , the results are less than the LOR and thus the RPD is not applicable.

pH in water Method: ME-(AU)-[ENV]AN101

Parameter	QC Reference	Units	LOR	MB	DUP %RPD	LCS %Recovery
pH**	LB017302	pH Units	0.1	5.7 - 6.1	0 - 5%	101%

Total and Volatile Suspended Solids (TSS / VSS) Method: ME-(AU)-[ENV]AN114

Parameter	QC Reference	Units	LOR	MB	DUP %RPD	LCS %Recovery
Total Suspended Solids Dried at 103-105°C	LB017314	mg/L	1	<1	1 - 3%	96 - 108%

METHOD

METHODOLOGY SUMMARY

AN083	Separatory funnels are used for aqueous samples and extracted by transferring an appropriate volume (mass) of liquid into a separatory funnel and adding 3 serial aliquots of dichloromethane. Samples receive a single extraction at pH 7 to recover base / neutral analytes and two extractions at pH < 2 to recover acidic analytes. QC samples are prepared by spiking organic free water with target analytes and extracting as per samples.
AN101	pH in Soil Sludge Sediment and Water: pH is measured electrometrically using a combination electrode (glass plus reference electrode) and is calibrated against 3 buffers purchased commercially. For soils, an extract with water is made at a ratio of 1:5 and the pH determined and reported on the extract. Reference APHA 4500-H+.
AN106	Conductivity and TDS by Calculation: Conductivity is measured by meter with temperature compensation and is calibrated against a standard solution of potassium chloride. Conductivity is generally reported as $\mu\text{mhos/cm}$ or $\mu\text{S/cm}$ @ 25°C. For soils, an extract with water is made at a ratio of 1:5 and the EC determined and reported on the extract, or calculated back to the as-received sample. Total Dissolved Salts can be estimated from conductivity using a conversion factor, which for natural waters, is in the range 0.55 to 0.75. SGS use 0.6. Reference APHA 2520 B.
AN114	Total Suspended and Volatile Suspended Solids: The sample is homogenised by shaking and a known volume is filtered through a pre-weighed GF/C filter paper and washed well with deionised water. The filter paper is dried and reweighed. The TSS is the residue retained by the filter per unit volume of sample. Reference APHA 2540 D. Internal Reference AN114
AN245	Anions by Ion Chromatography: A water sample is injected into an eluent stream that passes through the ion chromatographic system where the anions of interest ie Br, Cl, NO <sub>2</sub> , NO <sub>3</sub> and SO <sub>4</sub> are separated on their relative affinities for the active sites on the column packing material. Changes to the conductivity and the UV-visible absorbance of the eluent enable identification and quantitation of the anions based on their retention time and peak height or area. APHA 4110 B
AN277/WC250.312	Nitrite ions, when reacted with a reagent containing sulphanilamide and N-(1-naphthyl)-ethylenediamine dihydrochloride produce a highly coloured azo dye that is measured photometrically at 540nm.
AN279/AN293	The sample is digested with Sulphuric acid, K <sub>2</sub> SO <sub>4</sub> and CuSO <sub>4</sub> . All forms of phosphorus are converted into orthophosphate. The digest is cooled and placed on the discrete analyser for colorimetric analysis.
AN281	An unfiltered water or soil sample is first digested in a block digester with sulphuric acid, K <sub>2</sub> SO <sub>4</sub> and CuSO <sub>4</sub> . The ammonia produced following digestion is then measured colourimetrically using the Aquakem 250 Discrete Analyser. A portion of the digested sample is buffered to an alkaline pH, and interfering cations are complexed. The ammonia then reacts with salicylate and hypochlorite to give a blue colour whose absorbance is measured at 660nm and compared with calibration standards. This is proportional to the concentration of Total Kjeldahl Nitrogen in the original sample.
AN400	OC and OP Pesticides by GC-ECD: The determination of organochlorine (OC) and organophosphorus (OP) pesticides and polychlorinated biphenyls (PCBs) in soils, sludges and groundwater. (Based on USEPA methods 3510, 3550, 8140 and 8080.)
AN420	SVOC Compounds: Semi-Volatile Organic Compounds (SVOCs) including OC, OP, PCB, Herbicides, PAH, Phthalates and Speciated Phenols (etc) in soils, sediments and waters are determined by GCMS/ECD technique following appropriate solvent extraction process (Based on USEPA 3500C and 8270D).
AN420	SVOC Compounds: Semi-Volatile Organic Compounds (SVOCs) including OC, OP, PCB, Herbicides, PAH, Phthalates and Speciated Phenols in soils, sediments and waters are determined by GCMS/ECD technique following appropriate solvent extraction process (Based on USEPA 3500C and 8270D).



FOOTNOTES

IS	Insufficient sample for analysis.	LOR	Limit of Reporting
LNR	Sample listed, but not received.	↑↓	Raised or Lowered Limit of Reporting
*	This analysis is not covered by the scope of accreditation.	QFH	QC result is above the upper tolerance
**	Indicative data, theoretical holding time exceeded.	QFL	QC result is below the lower tolerance
^	Performed by outside laboratory.	-	The sample was not analysed for this analyte
		NVL	Not Validated

Samples analysed as received.  
Solid samples expressed on a dry weight basis.

Some totals may not appear to add up because the total is rounded after adding up the raw values.

The QC criteria are subject to internal review according to the SGS QAQC plan and may be provided on request or alternatively can be found here:  
<http://www.sgs.com.au/~media/Local/Australia/Documents/Technical%20Documents/MP-AU-ENV-QU-022%20QA%20QC%20Plan.pdf>

This document is issued, on the Client's behalf, by the Company under its General Conditions of Service available on request and accessible at <http://www.sgs.com/en/Terms-and-Conditions/General-Conditions-of-Services-English.aspx>. The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents.

This report must not be reproduced, except in full.

Order  
Project Mungindi Discharge Event Monitoring

Sample Name BE011457.001 BE011457.002 BE011457.003  
Description MG Site 3 Drainag MG Site 4 Sedime MG Site 5 Sedimentation Pond W Module Yard  
Sample Date 5/1/2015 5/1/2015 5/1/2015  
Matrix Water Water Water

Job Number	Method Name	Analyte Name	Units	Reporting Limit	Result	Result	Result
BE011457	pH in water	pH**	pH Unit	0.1	6.7	6.6	7.3
BE011457	Conductivity and TDS by Calcul	Conductivity @ 25 C	µS/cm	5	110	43	230
BE011457	Total and Volatile Suspended Sol	Total Suspended Sol	mg/L	1	380	720	120
BE011457	TKN Kjeldahl Digestion by Discr	Total Nitrogen (calc)	mg/L	0.05	1.5	1.8	2.1
BE011457	Total Phosphorus by Kjeldahl DI	Total Phosphorus (K	mg/L	0.01	0.65	0.74	0.86
BE011457	OC Pesticides in Water	Alpha BHC	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Hexachlorobenzene	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Beta BHC	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Lindane (gamma BH	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Delta BHC	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Heptachlor	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Aldrin	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Heptachlor epoxide	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Isodrin	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Gamma Chlordane	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Alpha Chlordane	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Alpha Endosulfan	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	p,p'-DDE	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Dieldrin	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Endrin	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Beta Endosulfan	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	p,p'-DDD	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Endosulfan sulphate	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	p,p'-DDT	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Endrin ketone	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Methoxychlor	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	Mirex	µg/L	0.1 <0.1	<0.1	<0.1	
BE011457	OC Pesticides in Water	d14-p-terphenyl (Sur %		0	78	60	66
BE011457	OC Pesticides in Water	2-fluorobiphenyl (Sur %		0	72	62	66
BE011457	OC Pesticides in Water	d5-nitrobenzene (Sur %		0	76	52	64
BE011457	OP Pesticides in Water	Dichlorvos	µg/L	1 <1	<1	<1	
BE011457	OP Pesticides in Water	Dimethoate	µg/L	1 <1	<1	<1	
BE011457	OP Pesticides in Water	Diazinon (Dimpylate)	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	OP Pesticides in Water	Fenitrothion	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Malathion	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Chlorpyrifos (Chlorp)	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Parathion-ethyl (Par	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Bromophos Ethyl	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Methidathion	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	OP Pesticides in Water	Ethion	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	Azinphos-methyl	µg/L	0.2 <0.2	<0.2	<0.2	
BE011457	OP Pesticides in Water	d14-p-terphenyl (Sur %		0	78	60	66
BE011457	OP Pesticides in Water	d5-nitrobenzene (Sur %		0	72	62	66
BE011457	OP Pesticides in Water	2-fluorobiphenyl (Sur %		0	76	52	64
BE011457	Other SVOC Analytes in Water	Carbophenothion	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Chlorpyrifos-methyl	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Dichlofenthion	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Dioxathion	µg/L	2 <2	<2	<2	
BE011457	Other SVOC Analytes in Water	Famphur (Famophos	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Fonophos	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Terbufos	µg/L	0.5 <0.5	<0.5	<0.5	
BE011457	Other SVOC Analytes in Water	Thionazin	µg/L	1 <1	<1	<1	